

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Measuring instrument Fourier Transform Infra Red (FTIR) Spectrometer
Name and address of the applicant	ABB Inc. 3400 Rue Pierre-Ardouin Quebec QC G1P 0B2 CANADA
Name and address of the manufacturer	ABB Inc. 3400 Rue Pierre-Ardouin, Quebec QC G1P 0B2, CANADA
Name and address of the factory	ABB Inc. 3400 Rue Pierre-Ardouin, Quebec QC G1P 0B2, CANADA
Ratings and principal characteristics	Option 1: Rated Input Voltage: +15 V DC/ -15 V DC/ 5 V DC Rated Input Current: 4 A/ 1 A/ 10A Option 2: Rated Input Voltage: 100-240 V AC Rated Frequency: 50/60 Hz Rated Input Power: 100 VA Degree of Protection: IP3X
Trademark / Brand (if any)	ABB
Model/type Ref.	MB 3000 and MB 3600
Additional information (if necessary)	Certificate DE 3 - 30385 issued 2017-02-24 is replaced by this version due to non technical change.
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016
as shown in the Test Report Ref. No. which forms part of this certificate	7169011428-000

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 064584 0051 Rev. 00

Date, 2022-07-19



(Timothee Djoto Kenfack)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service