



Ref. Certif. No.

DE 2-029400

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

**CB TEST CERTIFICATE**

Product	Switching mode power supply
Name and address of the applicant	ABB STOTZ-KONTAKT GmbH Eppelheimer Str. 82 69123 Heidelberg, Germany
Name and address of the manufacturer	ABB STOTZ-KONTAKT GmbH Eppelheimer Str. 82 69123 Heidelberg, Germany
Name and address of the factory	See additional page(s)
Ratings and principal characteristics	Input:100-240VAC, 50-60Hz, max.1.46A or 100-250VDC, max.1.0A Output: 24-28VDC, 72W (-25...+70°C, > 60°C derating 2.5%/K ) Power reserve current (5s): 4.5A (-25...+50°C) Relay output: 24Vdc, 1A; Class I
Trademark (if any)	ABB
Customer's Testing Facility (CTF) Stage used	N/A
Model / Type Ref.	CP-S.1 24/3.0
Additional information (if necessary may also be reported on page 2)	
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010+A1 for national differences see test report
As shown in the Test Report Ref. No. which forms part of this Certificate	60390275 001

This CB Test Certificate is issued by the National Certification Body



TÜV Rheinland LGA Products GmbH  
Tillystr. 2, 90431 Nürnberg, Germany  
Phone + 49 221 806-1371  
Fax + 49 221 806-3935  
Mail: cert-validity@de.tuv.com  
Web : www.tuv.com

Date: 2021-01-27

Signature: Dipl.-Ing. Univ. S. O. Steinke