



Manufacturer's Declaration

Manufacturer: **ABB AG (Measurement & Analytics)**
Address: **Schillerstraße 72
D-32425 Minden**
Document number: **2024-007**
Product name: **Coriolis Massflowmeters**
Type: **CoriolisMaster (FCB), CoriolisMaster
Hygienic (FCH) and Transmitter (FCT)
Series 400**

We hereby confirm that the following materials that are intended to come into contact with the fluid to be measured fulfill the requirements of the Code of Federal Regulations Title 21 (CFR 21) of the Food and Drug Administration (FDA):

Wetted parts made from AISI 316L / 1.4435 / 1.4404

CFR 21 Part 182*, Part 189*

Restrictions:

* Reviewing parts 182 (substances generally recognized as safe) and 189 (substances prohibited from use in human food) no statements against the use of stainless steel were found. Thus, we consider these materials as GRAS (generally recognized as safe). We will review our position in case of evidence that our evaluation is either incorrect or incomplete.

The information stated in this declaration reflects the state of the products according to valid datasheet at the time of document creation. Due to changes, deviations may occur later. To ensure that this declaration is also valid at a later point in time, the creator of this document must be contacted.

21st May 2024

Datum /
Date

i.V. Frank Gotthardt
Head of Development

i.V. Tomasz Pilch
QMS & OpEx Manager