

L150 Custom Configuration

1 Introduction

ABB can supply custom configurations for the L150 Level Indicator on request.

Enter the required setting or place a check mark (✓) against the relevant parameters in the following tables and return this document to the Global Sales office at Stonehouse.

2 Basic Hardware and Configuration

2.1 Hardware Assignment and Input Type

Referring to Section 4.3.1 of the User Guide (IM/L150), enter the setting required for each of the following parameters:

Frequency		Relay 1 Source	Relay 2 * Source	Relay 3 * Source	Logic Output Source	Analog Output Source	✓ Required Configuration
50 Hz	60 Hz						
<i>i</i>	<i>fl</i>	Alarm 1	Alarm 2	Alarm 3	Input Fail	Measured Level	
<i>z</i>	<i>b</i>	Alarm 1	Alarm 2	Alarm 3	Input Fail	Measured Level Average	
<i>U</i>		Custom	Custom	Custom	Custom	Custom	

* Available only if the appropriate option board is fitted

Continued...

Input Type Configuration (✓ the input type required)

4 to 20 mA	
1 to 10 mA (P851, P861 sensor)	
Custom	

Decimal Places (✓ the number of decimal places required)

0	
1	
2	
3	
4	

2.2 Alarms and Set Points

Referring to Section 4.3.2 of the User Guide (IM/L150), enter the setting required for each of the following parameters:

Alarm 1 Type (✓ the alarm type required)

None	
High Process	
Low Process	
High Latch	
Low Latch	
Fast Rate	
Slow Rate	

Alarm 1 Trip (enter the Alarm 1 trip point value)

Alarm 2 Type (✓ the alarm type required)

None	
High Process	
Low Process	
High Latch	
Low Latch	
Fast Rate	
Slow Rate	

Alarm 2 Trip (enter the Alarm 2 trip point value)

Alarm 3 Type (✓ the alarm type required)

None	
High Process	
Low Process	
High Latch	
Low Latch	
Fast Rate	
Slow Rate	

Alarm 3 Trip (enter the Alarm 3 trip point value)

Alarm Hysteresis (enter the Alarm hysteresis value)

2.3 Digital Input and Serial Communications

Referring to Section 4.3.4 of the User Guide (IM/L150), enter the setting required for each of the following parameters:

Digital Input Functions (✓ the function required)

None	
Average, Max./Min. Reset	
Front Panel Lockout	
Alarm Acknowledge	

Analog Input Filter (✓ the filter value required)

0 Seconds	
1 Second	
2 Seconds	
5 Seconds	
10 Seconds	
20 Seconds	
40 Seconds	
60 Seconds	

Serial Communication Configuration (✓ the configuration required)

Off	
2400 Baud, 2-wire	
2400 Baud, 4-wire	
9600 Baud, 2-wire	
9600 Baud, 4-wire	

Serial Communication Parity (✓ the parity required)

None	
Odd	
Even	

2.4 Ranges and Passwords

Referring to Section 4.4 of the User Guide (IM/L150), enter the value required for each of the following parameters:

Engineering Range (enter the values required)

High	
Low	

Retransmission Range (enter the values required)

High	
Low	

Volume Constant (enter the value required)

Set Up Password

(enter the password required between 0 and 9999)

Configuration Password

(enter the password required between 0 and 9999)

Modbus Address

(enter the address required between 0 and 99)

ABB has Sales & Customer Support expertise
in over 100 countries worldwide

www.abb.com

The Company's policy is one of continuous product
improvement and the right is reserved to modify the
information contained herein without notice.

Printed in UK (12.08)

© ABB 2008



ABB Limited
Howard Road, St. Neots
Cambridgeshire
PE19 8EU
UK
Tel: +44 (0)1480 475321
Fax: +44 (0)1480 217948

ABB Inc.
125 E. County Line Road
Warminster
PA 18974
USA
Tel:+1 215 674 6000
Fax:+1 215 674 7183